## Quadrathane™ ALC-95A

Thermoplastic Polyurethane Elastomer (PC Based)

Biomerics, LLC

General Information

## Message:

Quadrathane<sup>™</sup> ALC-95A is high performance aliphatic polycarbonate thermoplastic polyurethane. The polymer is naturally clear and supplied in small pellets for ease of processing. The material exhibits excellent mechanical properties, oxidative stability, biocompatibility, superior biostability in long term implantable devices, chemical resistance, non --yellowing during aging and softening at body temperature. The resin has consistent melt flow properties making it ideal for extrusion.

Quadrathane™, Quadraflex™, Quadraban™ and Quadraplast™ performance polymers are primarily used in life science and medical applications including vascular access devices, surgical supplies, respiratory devices, tracheotomy devices, and other medical applications. Typical end products include tubing, catheter parts, balloons, and various medical device components. These performance polymers are available in a variety of durometers, radiopacifiers, colors, and custom formulations.

| Antioxidation  Workability, good  Good liquidity  Good color stability  Good chemical resistance  Biocompatibility  aliphatic |  |   |  |
|---|--|---|--|
| Good liquidity Good color stability Good chemical resistance Biocompatibility   |  |   |  |
| Good color stability  Good chemical resistance  Biocompatibility  |  |   |  |
| Good chemical resistance Biocompatibility   |  |   |  |
| Biocompatibility  |  |   |  |
|   |  |   |  |
| alinhatic   |  |   |  |
| апрпацс   |  |   |  |
|   |  |   |  |
| Pipe fittings   |  |   |  |
| Human implant   |  |   |  |
| Surgical instruments  |  |   |  |
| Medical/nursing supplies  |  |   |  |
|   |  |   |  |
| Clear/transparent   |  |   |  |
| Particle  |  |   |  |
|   |  |   |  |
| Injection molding   |  |   |  |
| Nominal Value   | Unit   | Test Method   |  |
| 1.15  | g/cm³  | ASTM D792   |  |
| 7.5   | g/10 min   | ASTM D1238  |  |
| 0.60 - 1.0  | %  | ASTM D955   |  |
| Nominal Value   | Unit   | Test Method   |  |
| 95  |  | ASTM D2240  |  |
| Nominal Value   | Unit   | Test Method   |  |
| 34.5  | MPa  | ASTM D790   |  |
| Nominal Value   | Unit   | Test Method   |  |
| 6.21  | MPa  | ASTM D412   |  |
|   | Human implant Surgical instruments Medical/nursing supplies  Clear/transparent Particle Extrusion Injection molding  Nominal Value  1.15  7.5  0.60 - 1.0  Nominal Value  95  Nominal Value  34.5  Nominal Value | Human implant Surgical instruments Medical/nursing supplies  Clear/transparent Particle Extrusion Injection molding  Nominal Value  1.15  7.5  g/10 min  0.60 - 1.0  Nominal Value  Unit  95  Nominal Value  Unit  MPa  Nominal Value  Unit |  |

| Tensile Stress                      |                                       |                                   | ASTM D412 |
|-------------------------------------|---------------------------------------|-----------------------------------|-----------|
| 100% strain                         | 12.1                                  | MPa                               | ASTM D412 |
| 300% strain                         | 27.6                                  | MPa                               | ASTM D412 |
| Tensile Strength (Break)            | 41.4                                  | MPa                               | ASTM D412 |
| Tensile Elongation (Break)          | 450                                   | %                                 | ASTM D412 |
| Thermoset                           | Nominal Value                         | Unit                              |           |
| Post Cure Time (38°C)               | 6.0 - 10                              | hr                                |           |
| Injection                           | Nominal Value                         | Unit                              |           |
| Drying Temperature                  | 54.4                                  | °C                                |           |
| Drying Time                         | 4.0                                   | hr                                |           |
| Suggested Max Moisture              | < 3.0E-3                              | %                                 |           |
| Rear Temperature                    | 177                                   | °C                                |           |
| Front Temperature                   | 191                                   | °C                                |           |
| Nozzle Temperature                  | 196                                   | °C                                |           |
| Processing (Melt) Temp              | 204                                   | °C                                |           |
| Mold Temperature                    | 4.44 - 32.2                           | °C                                |           |
| Injection Rate                      | Slow                                  |                                   |           |
| Screw Compression Ratio             | 2.5:1.0 - 3.5:1.0                     |                                   |           |
| Injection instructions              |                                       |                                   |           |
| Injection Speed: 10 g/secCooling/Ho | old Time: Long, at least 50% of cycle | 20 to 60 secs depending on thickr | ness)     |
| Extrusion                           | Nominal Value                         | Unit                              |           |
| Drying Temperature                  | 54.4                                  | °C                                |           |
| Drying Time                         | 4.0                                   | hr                                |           |
| Suggested Max Moisture              | < 0.030                               | %                                 |           |
| Cylinder Zone 1 Temp.               | 171                                   | °C                                |           |
| Cylinder Zone 2 Temp.               | 182                                   | °C                                |           |
| Cylinder Zone 3 Temp.               | 188                                   | °C                                |           |
| Cylinder Zone 4 Temp.               | 193                                   | °C                                |           |
| Melt Temperature                    | 193                                   | °C                                |           |
| Die Temperature                     | 193 - 216                             | °C                                |           |
| Back Pressure                       | 6.89 - 12.4                           | MPa                               |           |

Screen Pack: 250 meshScrew Speed: Low sheer, 150 to 250 rpmWater Bath: 80 to 110°F

The information and data on this page are provided by manufacturers and document providers. SHANGHAI SUSHENG assumes no legal liability. It is strongly recommended to verify all technical data with material suppliers before final material selection. All rights belong to the original authors. If any infringement occurs, please contact us immediately.

## Recommended distributors for this material

## Susheng Import & Export Trading Co.,Ltd.

Tel: +86 21 5895 8519 Phone: +86 13424755533 Email: sales@su-jiao.com

Extrusion instructions

No. 215, Lianhe North Road, Fengxian District, Shanghai, China

